PLASON®

CLEANING OF THE FILTER AND COOLING SYSTEM OF THE DEVICE

TO CLEAN THE INTERNAL COOLING SYSTEM YOU NEED TO PREPARE:

- bowl for waste water,
- 0.5 l of Cillit (cleaning fluid against limescale, manufacturer: Reckitt Benckiser),
- 1l of distilled water,
- 1l of alcohol 50%.

DESCRIPTION OF THE CLEANING PROCESS:

1. Switch off the device.
2. Screw off the tip (manipulator).
3. Prepare the bowl for used cooling liquid. Switch on the green switch and wait until the fluid is pumped out of the device.
4. Switch off the device.
5. Screw on the tip (manipulator).
6. Unscrew the cap on cooling liquid opening. Fill 0.45 l of Cillit to the cooling liquid opening.
7. Switch on the device. Let the device in standby regime (keep the switch on, but DON'T press the START button) for 5 minutes.
8. Switch off the device.
9. Wait for 20 minutes.
10. Switch on the device. Let the device in standby regime (keep the switch on, but DON'T press the START button) for 2 minutes.
11. Switch off the device. Prepare the bowl for used cleaning liquid. Unscrew the tip (manipulator).
12. Switch on the green switch and wait until the fluid is pumped out of the device.
13. Turn off the device. Screw the tip (manipulator) on again. Fill the opening for cooling liquid by 0.5 l of distilled water.
14. Repeat points 10., 11., 12., 13.
15. Repeat Points 10., 11. and 12.
16. Turn off the device. Screw the tip (manipulator) on again. Fill the opening for cooling liquid with necessary amount of cooling liquid – 50% alcohol.
17. Let the opening for cooling liquid be open. Turn on the green switch and let the device be in standby mode for 3 minutes.
18. Screw on the cap on the cooling liquid opening.

THE DEVICE SHOULD NOW BE CLEANED AND IT IS READY FOR OPERATION.